

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-29 VOLUME (ml) 99 LOT No. 27345B EXP. DATE: 12/2019

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.4	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 56 Count at 45 minutes: 67 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609) 584-7677


 Joseph Zindulis, Ph.D.
 Production Manager
January 23, 2018

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-29 VOLUME (ml) 99 LOT No. 28009A EXP. DATE: 1/2020

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.4	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 61 Count at 45 minutes: 61 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609) 584-7677


 Joseph Zindulis, Ph.D.
 Production Manager
February 28, 2018

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-29 VOLUME (ml) 99 LOT No. 28009B EXP. DATE: 1/2020

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.2	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 57 Count at 45 minutes: 60 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
 Weber Scientific
 2732 Kuser Road
 Hamilton, NJ 08691
 (609) 584-7677



 Joseph Zindulis, Ph.D.
 Production Manager

 February 28, 2018

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-29 VOLUME (ml) 99 LOT No. 28010A EXP. DATE: 1/2020

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.3	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 63 Count at 45 minutes: 53 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609) 584-7677


 Joseph Zindulis, Ph.D.
 Production Manager
February 28, 2018

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-29 VOLUME (ml) 99 LOT No. 28010B EXP. DATE: 1/2020

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.2	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 69 Count at 45 minutes: 59 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
 Weber Scientific
 2732 Kuser Road
 Hamilton, NJ 08691
 (609) 584-7677


 Joseph Zindulis, Ph.D.
 Production Manager
February 28, 2018

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-29 VOLUME (ml) 99 LOT No. 28023A EXP. DATE: 1/2020

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.2	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 84 Count at 45 minutes: 71 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609) 584-7677


 Joseph Zindulis, Ph.D.
 Production Manager
 March 8, 2018

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-29 VOLUME (ml) 99 LOT No. 28023B EXP. DATE: 1/2020

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.2	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 70 Count at 45 minutes: 71 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
 Weber Scientific
 2732 Kuser Road
 Hamilton, NJ 08691
 (609) 584-7677



 Joseph Zindulis, Ph.D.
 Production Manager

 March 8, 2018

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-29 VOLUME (ml) 99 LOT No. 28037A EXP. DATE: 2/2020

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.3	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 62 Count at 45 minutes: 63 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
 Weber Scientific
 2732 Kuser Road
 Hamilton, NJ 08691
 (609) 584-7677


 Joseph Zindulis, Ph.D.
 Production Manager
 March 28, 2018

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-29 VOLUME (ml) 99 LOT No. 28045A EXP. DATE: 2/2020

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.1	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 69 Count at 45 minutes: 71 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609) 584-7677


 Joseph Zindulis, Ph.D.
 Production Manager
April 10, 2018

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-29 VOLUME (ml) 99 LOT No. 28051A EXP. DATE: 2/2020

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.2	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 58 Count at 45 minutes: 65 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609) 584-7677


 Joseph Zindulis, Ph.D.
 Production Manager
April 12, 2018

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-29 VOLUME (ml) 99 LOT No. 28075A EXP. DATE: 3/2020

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.1	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 62 Count at 45 minutes: 77 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609) 584-7677


 Joseph Zindulis, Ph.D.
 Production Manager
April 23, 2018

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-29 VOLUME (ml) 99 LOT No. 28075B EXP. DATE: 3/2020

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.2	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 71 Count at 45 minutes: 76 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609) 584-7677


 Joseph Zindulis, Ph.D.
 Production Manager
April 23, 2018

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-29 VOLUME (ml) 99 LOT No. 28094B EXP. DATE: 4/2020

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.2	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 75 Count at 45 minutes: 75 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609) 584-7677


 Joseph Zindulis, Ph.D.
 Production Manager
May 24, 2018

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-29 VOLUME (ml) 99 LOT No. 28099B EXP. DATE: 4/2020

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.2	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 72 Count at 45 minutes: 67 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609) 584-7677


 Joseph Zindulis, Ph.D.
 Production Manager
 June 5, 2018

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-29 VOLUME (ml) 99 LOT No. 28128C EXP. DATE: 5/2020

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.2	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 68 Count at 45 minutes: 74 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609) 584-7677


 Joseph Zindulis, Ph.D.
 Production Manager
 June 29, 2018

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-29 VOLUME (ml) 99 LOT No. 28171C EXP. DATE: 6/2020

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.2	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 65 Count at 45 minutes: 60 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609) 584-7677


 Joseph Zindulis, Ph.D.
 Production Manager
August 13, 2018

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-29 VOLUME (ml) 99 LOT No. 28206C EXP. DATE: 7/2020

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.0	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 69 Count at 45 minutes: 59 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609) 584-7677


 Joseph Zindulis, Ph.D.
 Production Manager
September 14, 2018

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-29 VOLUME (ml) 99 LOT No. 28225C EXP. DATE: 8/2020

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.3	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 48 Count at 45 minutes: 57 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609) 584-7677


 Joseph Zindulis, Ph.D.
 Production Manager
September 18, 2018