

bioavid Diagnostics

Lateral Flow Principles / Standards

Sample Preparation Food Products (except chocolate)

For Nuts; Nut mixtures and Bakery

Weigh 50 g of the sample
Weigh 4 g salt

Pour into a mixer and add 450 ml dist. water
Homogenize for 5 min

Filter, centrifuge or allow homogenate to settle

From the aqueous phase add 0.1 ml, corresp. to 2 drops into the assay

Total time for sample preparation: less than 10 min

