




## CERTIFICATE OF ANALYSIS

PRODUCT DESCRIPTION: TwinSampler with DE-NB Buffer

CATALOG No. 3070-12 LOT No. 129155 EXP. DATE: 06/2021

TEST	RESULTS	SPECIFICATION
SPONGE HYDRATION	CONFORMS	DE NEUTRALIZING BROTH WITHOUT BROMOCRESOL PURPLE
pH (25° C)	MEASUREMENT	7.6 +/- 0.2
STERILITY	PASSES	IRRADIATED NO GROWTH
FILL VOLUME	CONFORMS	10 ml +/- 0.2 ml
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON
PACKAGE INTEGRITY INDIVIDUAL DEVICE	CONFORMS	NO LEAKAGE
INTEGRITY OF POUCH WITH 20 DEVICES	CONFORMS	INTACT SEAL & NO VISIBLE DAMAGE TO POUCH
PACKAGE LABELS	CONFORMS	CONFORMS TO STANDARD

Made By:  
Weber Scientific  
2732 Kuser Road  
Hamilton, NJ 08691  
(609) 584-7677

  
Joseph Zindulis, Ph.D.

Production Manager

July 17, 2019