



CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION Weber Scientific Phosphate Buffer with MgCl₂ CATALOG No. 3127-59
VOLUME (ml) 90 LOT No. 54316A EXP. DATE 11/2016 COLOR CODE Gold

TEST	RESULTS	SPECIFICATION	METHOD
pH @25° C	7.4	7.2 +/- 0.2	WS cSOP 8.2
TOXICITY*	Count at 0 minutes: 60 Count at 45minutes:62 PASSES	< 20% DECREASE	WS cSOP 24.3
STERILITY	PASSES	$<1 \times 10^{-6}$	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.2; 13.2
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.0
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.1
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.1
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.3
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.3; 6.2; 38.0

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609)-584-7677


Fred Weber
President

Date December 17, 2014



CERTIFICATE OF ANALYSIS

PRODUCT DESCRIPTION Weber Scientific Phosphate Buffer with MgCl₂ CATALOG No. 3127-59
VOLUME (ml) 90 LOT No. 54010A EXP. DATE 01/2016 COLOR CODE Gold

TEST	RESULTS	SPECIFICATION	METHOD
pH @25° C	7.31	7.2 +/- 0.2	WS cSOP 8.2
TOXICITY*	Count at 0 minutes: 54 Count at 45minutes:62 PASSES	< 20% DECREASE	WS cSOP 24.3
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.2; 13.2
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.0
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.1
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.1
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.3
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.3; 6.2; 38.0

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609)-584-7677


Fred Weber
President

Date February 7, 2014



CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION Weber Scientific Phosphate Buffer with MgCl₂ CATALOG No. 3127-59
VOLUME (ml) 90 LOT No.: 55135B EXP. DATE: 05/2017 COLOR CODE: Gold

TEST	RESULTS	SPECIFICATION	METHOD
pH @25° C	7.3	7.2 +/- 0.2	WS cSOP 8.2
TOXICITY*	Count at 0 minutes: 73 Count at 45minutes: 72 PASSES	< 20% DECREASE	WS cSOP 24.3
STERILITY	PASSES	$<1 \times 10^{-6}$	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.2; 13.2
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.0
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.1
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.1
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.3
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.3; 6.2; 38.0

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609)-584-7677


Fred Weber
President

Date: June 15, 2015



CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION Weber Scientific Phosphate Buffer with MgCl₂ CATALOG No. 3127-59
VOLUME (ml) 90 LOT No. 54339B EXP. DATE 12/2016 COLOR CODE Gold

TEST	RESULTS	SPECIFICATION	METHOD
pH @25° C	7.3	7.2 +/- 0.2	WS cSOP 8.2
TOXICITY*	Count at 0 minutes: 72 Count at 45minutes:74 PASSES	< 20% DECREASE	WS cSOP 24.3
STERILITY	PASSES	$<1 \times 10^{-6}$	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.2; 13.2
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.0
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.1
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.1
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.3
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.3; 6.2; 38.0

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609)-584-7677


Fred Weber
President

Date January 9, 2015



CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
CATALOG No. 3127-59 VOLUME (ml) 90 LOT No. 56020B EXP. DATE: 01/2018

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.2	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 59 Count at 45 minutes: 70 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	$<1 \times 10^{-6}$	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609) 584-7677


Fred Weber, President

Date: February 15, 2016

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-59 VOLUME (ml) 90 LOT No. 57033B EXP. DATE: 2/2019

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.2	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 61 Count at 45 minutes: 57 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609) 584-7677


 Joseph Zindulis, Ph.D.
 Production Manager
 March 29, 2017

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-59 VOLUME (ml) 90 LOT No. 58044A EXP. DATE: 2/2020

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.3	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 78 Count at 45 minutes: 80 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609) 584-7677


 Joseph Zindulis, Ph.D.
 Production Manager
 April 10, 2018

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-59 VOLUME (ml) 90 LOT No. 58205C EXP. DATE: 7/2020

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.1	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 75 Count at 45 minutes: 65 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609) 584-7677


 Joseph Zindulis, Ph.D.
 Production Manager
September 14, 2018

CERTIFICATE OF ANALYSIS


PRODUCT DESCRIPTION: Weber Scientific Phosphate Butterfield Buffer with MgCl₂
 CATALOG No. 3127-59 VOLUME (ml) 90 LOT No. 59232C EXP. DATE: 08/2021

TEST	RESULTS	SPECIFICATION	METHOD
pH @ 25° C	7.1	7.2 +/- 0.2	WS cSOP 8.4
TOXICITY*	Count at 0 minutes: 79 Count at 45 minutes: 75 PASSES	< 20% DECREASE	WS cSOP 24.3; 54.0
STERILITY	PASSES	<1 x10 ⁻⁶	AAMI** VALIDATED Gamma Irradiated (Certificate)
FILL VOLUME	CONFORMS	LABEL +/- 2 ml	WS cSOP 9.4; 13.4
COLOR	CONFORMS	CLEAR NO FOREIGN MATTER	VISUAL WS cSOP 33.2
PACKAGE INTEGRITY SHIPPER	CONFORMS	NO VISIBLE DAMAGE TO CARTON	VISUAL WS cSOP 15.3
PACKAGE INTEGRITY BOTTLE	CONFORMS	NO LEAKING	WS cSOP 14.4
PACKAGE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5
BOTTLE MARKINGS	CONFORMS	CONFORMS TO STANDARD	WS cSOP 5.5; 6.4; 38.2

*TEST PERFORMED AT COMMERCIAL LAB: New Jersey Laboratories.

**Association for the Advancement of Medical Instrumentation validation method ANSI/AAMI/ISO 11137-1994; microbiological testing method AAMI TIR 8-1991.

Made By:
Weber Scientific
2732 Kuser Road
Hamilton, NJ 08691
(609) 584-7677


 Joseph Zindulis, Ph.D.
 Production Manager
October 3, 2019